

湖北省计量测试技术研究院

Huber Institute of Measurement and Testing Technology

地址 (Add) : 湖北省武汉市东湖新技术开发区茅
店山中路二号 邮编 (Post Code) : 430223
网址 (Web site) : <http://www.hbjl.gov.cn> 电话 (Tel) : 027-81925136
传真 (Fax) : 027-81925137

中国校准服务联合体
China United Calibration Service

CUC

REGISTRATION NO. 002

校准证书 CALIBRATION CERTIFICATE

证书编号:

Certificate No.

委托方名称 Customer	江苏省淮安水利枢纽维修养护中心(江苏江海水利水 电工程检测有限公司)
委托方地址 Address	江苏省淮安市淮安区漕运镇(原三堡乡)省灌溉总渠管 理处维修养护中心
样品名称 Name Of Sample	变压器组别测试仪
制造厂商 Manufacturer	武汉特高压电力科技有限公司
型号规格 Model/Type	UHV-332
器具编号 No Of Sample	202061

证书专用章

Stamp



校准日期
Calibration date

2022 年 07 月 26 日
Y M D

证书批准人
Approved by

耿睿

核验员
Checked by

李亮波

校准员
Calibrated by

董妙宇

本次校准所使用的测量装置均溯源至保存在中国计量科学研究院的国家计量基准。中国计量科学研究院于1999年代表中国签署了国际间“国家计量基标准及国家计量研究院出具的校准和测量证书相互承认协议”。

The measuring equipment used in the calibration is traceable to national primary standards maintained in National Institute of Metrology (NIM). NIM is the signatory to the Mutual Recognition Arrangement (MRA) for national measurement standards and for calibration and measurement certificates issued by national metrology institutes.



- 本院是政府计量行政管理部门依法设立的法定计量检定机构

This laboratory is a legal metrological verification institution established by the government metrological administrative department according to law.

- 本院质量管理体系符合ISO/IEC17025标准的要求。

The quality management system for laboratory complies with ISO/IEC 17025 standards.

- 本次校准的技术依据（名称、代号）

Reference documents for the Calibration (Name , Code)

参照: JJG 970—2002 变压比电桥检定规程 Verification Regulation of Transformers Turn Ratio Test Sets

- 本次校准所使用的主要计量标准器具

Main standards of measurement used in the Calibration

设备名称
Name of Equipment
感应分压器

型号/编号
Model/Serial No.
FG2/7216

证书号/有效期
Certificate No./Due Date
(计)字
第202211747号/2023-07-31

- 校准环境条件

Environmental condition on the Calibration

温度: 20.1 °C

相对湿度: 50 %

其它:

Temperature

R.H.

Others

气压:

地点: 光谷基地B132

原始记录编号: XXXXXXXXXX

Pressure

Place

Record No.

本校准结论, 仅对受校样品的本次校准有效。

It's Effect That Results of This Report Relate Only To The Sample(s) Calibrated.

未经本院许可, 不得部分复制本证书。

校准数据/结果

Data/Results of Calibration

变比测量:		
标准变比值	实测变比值	测量不确定度 $U, k=2$
5	4.9988	0.0060
10	9.9988	0.0090
15	15.000	0.015
30	30.010	0.030
50	50.055	0.050
100	100.205	0.090
300	299.98	0.30
500	499.54	0.50
1000	998.8	1.0

以下空白

研究院
章